



LumiraDx™ SARS-CoV-2 Antigen (Ag) Test Specifications

For Emergency Use Authorization Only (EUA). For *in vitro* Diagnostic Use. Rx Only.

Intended Use*

The LumiraDx SARS-CoV-2 Ag Test is a rapid microfluidic immunofluorescence assay for use with the LumiraDx Platform intended for the qualitative detection of the nucleocapsid protein antigen to SARS-CoV-2 directly from nasal swab specimens collected from individuals suspected of COVID-19 by their healthcare provider within the first twelve days of symptom onset.

Testing is limited to laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C 263a, that meet the requirements to perform moderate, high or waived complexity tests. This Test is authorized for use at the Point of Care (POC), i.e., in patient care settings operating under a CLIA Certificate of Waiver, Certificate of Compliance, or Certificate of Accreditation.

Test Description

The LumiraDx SARS-CoV-2 Ag Test uses SARS-CoV/SARS-CoV-2 specific antibodies in a particle-particle sandwich immunoassay to determine the presence of SARS-CoV-2 Nucleocapsid Protein (NP) antigen present in the test sample.

Built-in Quality Controls

The LumiraDx Platform Instrument and Test Strip are integrated with several control checks to ensure the Instrument and Test are functioning correctly for every test run. These checks include:

- Electrical component operation, heater operation, battery charge state, mechanical actuators and sensors and optical system performance
- Test Strip positioning, optics, and Test Strip expiry
- Monitoring of Test Strip performance and controls during test runtime
- The SARS-CoV-2 Ag Test contains an Onboard Quality Control (OBC) assay

SARS-CoV-2 Ag External Quality Controls

Positive and Negative Quality Controls are available from LumiraDx to complete Quality Control assessment of the Instrument and SARS-CoV-2 Ag Test Strips.

Clinical performance

Direct nasal swabs (257) were prospectively collected from symptomatic patients suspected of COVID-19 from six sites across the United States and United Kingdom. The performance of the LumiraDx SARS-CoV-2 Ag Test was compared to an EUA authorized PCR method.

LumiraDx SARS-CoV-2 Ag results	Reference PCR results		
	POS	NEG	Total
POS	81	6	87
NEG	2	168	170
Total	83	174	257

PPA: 97.6% (CI 91.6%- 99.3%)

NPA: 96.6% (CI 92.7%- 98.4%)

OPA: 96.9% (CI 94.0%- 98.4%)

PPA - Positive Percent Agreement; NPA - Negative Percent Agreement; OPA - Overall Percent Agreement, CI - Confidence Interval

Analytical performance

Limit of Detection

Starting material concentration	Estimated LoD	No. Positive/Total	% Positive
2.8 x 10 ⁵ TCID ₅₀ /mL	32 TCID ₅₀ /mL	20/20	100

*See SARS-CoV-2 Ag Test Product Insert for full Intended Use statement.

Cross reactivity

SARS-CoV-2 Ag Test was found not to cross-react with a panel of organisms and viruses including several human coronaviruses. See LumiraDx SARS-CoV-2 Ag Test Product Insert for full details.

Specifications

Sample type	Nasal swab
Time to result	12 minutes
Result display	Qualitative – positive or negative
Storage temperature	2-30 °C (36-86 °F)
Operating temperature	15-30 °C (59-86 °F)
Interferences	See LumiraDx SARS-CoV-2 Ag Test Product Insert for details
Onboard control	Onboard Quality Control (OBC) assay and sample processing control
Quality control material	Positive and Negative external liquid controls

Swabs

When collecting nasal swab samples, the following swabs have been validated for use with the LumiraDx SARS-CoV-2 Ag Test: Copan Nasal FLOQswab™ Regular, Puritan HydraFlock™ Sterile Standard Flock Swab, Aspen Surgical™ Polyester Swab, SteriPack™ Sterile Polyester Spun Swab, mwe medical wire Dryswab™ Rayon Swab, Kang Jian™ Virus Collection Swab. Swabs from other suppliers have not been validated and the performance with other swabs may not perform as expected. This is a complete list of swabs currently validated for use with the LumiraDx SARS-CoV-2 Ag Test.

Commercial availability of swabs may vary by country. Please be aware that the CDC does not recommend use of calcium alginate swabs or swabs with wooden shafts, as they may contain substances that inactivate some viruses and inhibit PCR testing.¹

For more information visit lumiradx.com or contact the LumiraDx Customer Services by email: CustomerServices.US@lumiradx.com or Tel: 1-888-586-4721

Copyright © 2020 LumiraDx UK LTD. All rights reserved worldwide.

LumiraDx and Flame logo are trademarks of LumiraDx International LTD. Full details of these and other registrations of LumiraDx can be found at lumiradx.com/IP. All other trademarks are the property of their respective owners.

Content should be used for the use of the LumiraDx products only and in line with instructions provided. You may not, except with our express written permission, distribute or commercially exploit the content. Nor may you transmit it or store it in any other form of electronic retrieval system other than for the purpose of use of the LumiraDx Instrument or LumiraDx Test Strips. Information provided is subject to change without notice.

Product is not available in all countries and regions. Please check with your local LumiraDx sales representative or distributor for availability in specific markets. Available in the USA under FDA Emergency Use Authorization.

In the USA, this test has not been FDA cleared or approved; this test has been authorized by FDA under an EUA for use by authorized laboratories; use by laboratories certified under the CLIA, 42 U.S.C. §263a, that meet requirements to perform moderate, high or waived complexity tests. This test is authorized for use at the Point of Care (POC), i.e., in patient care settings operating under a CLIA Certificate of Waiver, Certificate of Compliance, or Certificate of Accreditation. This test has been authorized only for the detection of proteins from SARS-CoV-2, not for any other viruses or pathogens. In the USA, this test is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of the virus that causes COVID-19 under Section 564(b)(1) of the Federal Food, Drug, and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the authorization is terminated or revoked sooner.

lumiradx.com

S-COM-ART-00458 R1 Date of Rev 2020/09

¹see <https://www.cdc.gov/coronavirus/2019-ncov/lab/guidelines-clinical-specimens.html>



Manufactured by:
LumiraDx UK Ltd
Dumyat Business Park
Alloa
FK10 2PB, UK

Registration Number:
09206123

US Office:
LumiraDx
221 Crescent Street
Suite 502
Waltham
Massachusetts 02453
USA